



# DILLINGER HÜTTE Laboratory

Laboratory accredited by GAZ, (Association for the Accreditation and Certification GmbH) Attestation no. 91021

## CERTIFIED REFERENCE MATERIAL SE21-56 unalloyed steel



Version: 002    Date of certification: 17.07.08    Sample form: Milling 0.200-1.400 mm

Analyte	Unit	Certified value	Confidence interval (95%)	No. of independent repeated series	Measuring method*
C	ppm	48.3	0.8	12	1
S	ppm	151.5	1.2	12	2

\* Measuring method

1: Combustion with coulometric detection

2: Combustion with IR-detection